

Establishment Name	Insp. Type	Insp. Date	Fail?	Rule Cited	Description of Violation	Severity
NEAL, MARK A JR	REGULAR	2019-09-04	No	3.E.1	Floors not clean, in good repair and/or all surfaces not easily cleanable and smooth.	N
				5.B.4	Autoclave not spore tested at least once each month, packages used for sterilizing instruments do not have chemical/heat sensitive indicators, sterilized instruments not re-sterilized after six months and/or spore test records not retained and available for inspection for a minimum of one year.	C
NEAL, MARK A JR	REGULAR	2021-05-04	No	3.E.1	Floors not clean, in good repair and/or all surfaces not easily cleanable and smooth.	N