

Establishment Name	Insp. Type	Insp. Date	Fail?	Rule Cited	Description of Violation	Severity
JOHNSON, ANDREW	REGULAR	2019-03-12	No	5.B.4	Autoclave not spore tested at least once each month, packages used for sterilizing instruments do not have chemical/heat sensitive indicators, sterilized instruments not re-sterilized after six months and/or spore test records not retained and available for inspection for a minimum of one year.	C