

Establishment Name	Insp. Type	Insp. Date	Fail?	Rule Cited	Description of Violation	Severity
ARTHURS, DAMIEN N	REGULAR	2020-12-11	No	4.S	Tattoo practitioner did not provide proper written and verbal aftercare instructions to the client after a tattoo procedure.	C
				5.B.4	Autoclave not spore tested at least once each month, packages used for sterilizing instruments do not have chemical/heat sensitive indicators, sterilized instruments not re-sterilized after six months and/or spore test records not retained and available for inspection for a minimum of one year.	C
				6.C	Patron records not retained and made available for at least two (2) years and/or a receipt with required information was not issued to a client at time of tattoo procedure visit.	N
ARTHURS, DAMIEN N	REGULAR	2023-11-07	No	5.B.4	Autoclave not spore tested at least once each month, packages used for sterilizing instruments do not have chemical/heat sensitive indicators, sterilized instruments not re-sterilized after six months and/or spore test records not retained and available for inspection for a minimum of one year.	C
				6.C	Patron records not retained and made available for at least two (2) years and/or a receipt with required information was not issued to a client at time of tattoo procedure visit.	N