

Establishment Name	Insp. Type	Insp. Date	Fail?	Rule Cited	Description of Violation	Severity
SCAMMON. BROOKE	REGULAR	2019-03-25	No	3.F.2	Proper lavatory facilities, anti-bacterial hand cleanser and/or disposable sanitary towels not provided at each workstation	C
				5.B.4	The sterilizer is not spore tested at least once each month, packaged sterilized instruments not monitored for sterilization by the use of chemical/heat sensitive indicators, packaged sterilized instruments not re-sterilized after 6 months before use, spore test records not maintained for one year and/or sterilizer spore test records not available for inspection.	C