

Establishment Name	Insp. Type	Insp. Date	Fail?	Rule Cited	Description of Violation	Severity
BOWLEY, ELAINE M	REGULAR	2022-07-19	No	5.B.4	The sterilizer is not spore tested at least once each month, packaged sterilized instruments not monitored for sterilization by the use of chemical/heat sensitive indicators, packaged sterilized instruments not re-sterilized after 6 months before use, spore test records not maintained for one year and/or sterilizer spore test records not available for inspection.	C