

| Establishment Name | Insp. Type | Insp. Date | Fail? | Rule Cited | Description of Violation  | Severity |
|--------------------|------------|------------|-------|------------|---|----------|
| BOWLEY, ELAINE M   | REGULAR    | 2022-07-19 | No    | 5.B.4      | The sterilizer is not spore tested at least once each month, packaged sterilized instruments not monitored for sterilization by the use of chemical/heat sensitive indicators, packaged sterilized instruments not re-sterilized after 6 months before use, spore test records not maintained for one year and/or sterilizer spore test records not available for inspection. | C        |